

For more information, please contact:

Distributor

NextDent B.V.
Centurionbaan 190
3769 AV Soesterberg
The Netherlands

+31 88 616 04 40
info@nextdent.com
www.nextdent.com

Manufacturer

Vertex-Dental B.V.
Centurionbaan 190
3769 AV Soesterberg
The Netherlands

+31 88 616 04 40
info@vertex-dental.com
www.vertex-dental.com

NextDent®

www.nextdent.com

NextDent, the leading manufacturer of dental materials for 3D printing.

NextDent B.V. is an independent subsidiary company under the Vertex Global Holding. NextDent's mission is to become the worldwide leading manufacturer of CE-certified and biocompatible dental 3D printing materials. NextDent's Research and Development team is constantly searching for the best possible solutions in order to become a benchmark for 3D printing materials in the dental field. The company's focus is on development of custom-made 3D printing solutions in close cooperation with our customers.

© 2016 Vertex-Dental B.V. All Rights Reserved

NextDent® is a registered trademark of Vertex-Dental B.V. used under a temporary license granted by NextDent B.V.. Neither, NextDent B.V. nor any of its affiliates is responsible for the manufacturing of the product(s).

BASE

Monomer based on acrylic esters for manufacturing
of 3D-printed denture bases



NextDent®

Leading manufacturer of dental materials for 3D printing

Instruction For Use

Base

NextDent™ Base is a monomer based on acrylic esters
for manufacturing of 3D-printed denture bases.
NextDent™ Base is a class IIa material and CE-certified.

The following instructions for use are for dental professionals who use NextDent™ Base as a denture base material for dentures. NextDent™ Base is intended exclusively for professional dental work. This instruction for use provides also information about safety and environmental aspects, a safety datasheet is available on www.nextdent.com and at local dealers. In case more information is needed about the processing of NextDent™ Base material contact the NextDent office. See also information at the end of this document.

NextDent™ Base is a monomer based on acrylic esters for manufacturing of 3D-printed denture bases, normally intended for permanent use. NextDent™ Base is a class IIa material and CE-certified.

IFU ID: INDB201703UK

Description and effects

NextDent™ Base can be used in combination with all laser and DLP based 3D printers which support NextDent materials.

Contra-indication

NextDent™ Base should not be used for any other purpose than denture base work only. Any deviation from this instruction for use may have an adverse on the chemical and physical quality of NextDent™ Base. In case of an allergic reaction, please contact a medical physician.

Hazard & Precaution (H&P phrases)

Inhalation:
Irritating to respiratory system. High atmospheric concentrations may lead to irritation of the respiratory tract, dizziness, headache and anesthetic effects.

Skin contact:
May cause sensitization by skin contact. Irritating to skin, repeated and/or prolonged contact may cause dermatitis.

Eye contact:
High vapor concentration may cause irritation.

Ingestion:
Low oral toxicity, but ingestion may cause irritation of the gastrointestinal tract.

Protection:
Wear protection when handling NextDent™ Base. Protective glasses and gloves are advised. Information about the handling of the product can be found in the safety datasheet, which is available on www.nextdent.com.

Hazard Phrases:
Pictogram



Signal word Warning

Hazard statements
H317 May cause an allergic skin reaction
H413 May cause long lasting harmful effect to aquatic life

Processing / Post-curing

Make sure that you work as clean as possible, dirty reservoirs or machines can cause deformation and therefore failure of the printed objects!




Thoroughly mix before using
Before using the material for the first time. Mix it for 2.5 hours in its original packaging. Each subsequent time you use this material. Mix for 1 hour. We advise to use a roller bench to thoroughly mix the material. Color deviation and print failures may occur when mixed insufficiently.



Fill printerreservoir
Pour the liquid material in the reservoir of the 3D-printing machine.



For printer settings see IFU of 3D Printer
Follow the instructions for use of the printer. These are delivered together with the pinter.



Remove printed parts from platform
When the machine has finished its program remove the building platform from the machine. Place the platform on some paper or cloth with the built jobs facing upwards. The printed jobs can know be removed from the platform using a suitable knife.



Cleaning Pieces step 1
Rinse the printed jobs for three minutes in an alcohol solution (>90%) to eliminate any excess material, using of an ultrasonic bath.



Cleaning Pieces step 2
Rinse for two minutes in a clean alcohol solution (>90%). Rinsing in a alcohol solution should not take longer than 5 minutes, as this may cause defects in the printed parts.



Post-Cure
After cleaning make sure the printed parts are dry and free of alcohol residual. Then place the printed jobs in a UV- light curing box for final polymerization.

Post-curing is an UV-light treatment to ensure that NextDent materials obtain full polymer conversion. Through this the residual monomer is reduced to a minimum and the required mechanical properties are obtained. This procedure is a necessary step to produce a biocompatible end-product. We strongly advice to make use of the NextDent™ LC 3D Printbox.

Specific curing-time NextDent™ Base

NextDent Material	Time (min.)	Wavelength (nm.)	Total output Light (Watt)	UV lightbox output W*Sek=J (kJ)
Base	30	Blue UV-A 315-400 UV-Blue 400-550	UV-A 108 UV-Blue 108	388.8

The unit used at NextDent has 6x 18W/71 lamps (Dulux L Blue) and 6, 18W/78 lamps (Dulux blue UV-A). The calculated output is based on the UV light UVA lamp Blue. Please notice that the light sources and the printing machine need a routine maintenance following the manufacturer instructions.

Finishing
Remove any support structures and finish jobs if necessary, using conventional dental methods and instruments. Differences in color nuance may occur due to production in batches of the raw material and product, inadequate shaking of the original packaging before use or insufficient post-curing.

Storage conditions, expiry date and transport

Store the product in the original packaging at room-temperature in a dry and dark area, preferably not exceeding 25°C. Close the packaging after each use. The expiry date of the product is mentioned on the product label. In case of exceeding the expiry date, the product is no longer guaranteed in terms of treatment. Do not expose to UV-light and moisture.

Plastic and packaging waste

The product NextDent™ Base in its polymerized form is not environmentally harmful. Residual waste material in its liquid state should be delivered to a collection point for chemical waste material.

Cleaning instructions

NextDent 3D-printing material should be cleaned with non-chemical products. If disinfecting before intended use is required, an ethanol solution can be used.

Delivery units

The product NextDent™ Base is available in the following packaging size: 1000 gr.

Explanation of symbols on labelling

- CE mark with NB number
- Batch number of product
- Manufacturer
- Keep away from sunlight
- Consult instructions for use
- Use-by date