## **DOC - DECLARATION OF CONFORMITY** € everall7 **QUALITY SYSTEM FORM DOCUMENT NUMBER** EA7-DOC-KW-AK05/V150/EN REVISION LEVEL PAGE NO Page 1 of 1

## DECLARATION OF CONFORMITY IN ACCORDANCE WITH REGULATION (EU) 2017/74 No: KW-AK05/V150/EN

Manufacturer:  Authorized EU-representative:		Everall7 Sp. z o.o.  ul. Augustówka 14 02-981 Warsaw (Poland) Single registration number (SRN): PL-MF-000018646  Name of the site, Address of the site – N/A				
						PRODUCT CODE / CATALOGUE NUMBER
V150Z03	Villacryl So	oft 40 ml Liquid + 60 g Powder + 10 ml Varnish		5		
V150L02		/illacryl Soft Liquid 40 ml				
V150P01		Villacryl Soft Varnish 10 ml		5		
Intended use of the product:	Acrylic resi	n for soft denture relining				
* Classification rules		ACCORDING TO ANNEX VIII OF THE REGULATION 2017/745	1271			
Basic UDI-DI		590160212RELFF				
DEVICES, IN ACCORDANCE WITH THE ODCUMENTATION IS RETAINED UNDER TH		EN ISO 13485 :2016 EN ISO 14971 :2019 EN ISO 20417:2021 EN ISO 15223-1:2021 EN ISO 10993-1: 2020 EN ISO 10993-5:2009				
		EN ISO 10993-10:2013 EN ISO 10993-18:2020				
Technical standards ref:		FACTORY STANDARD ZN-19/10 - Villacryl Soft – Acrylic material for soft relining of dentures				
EC certificate:		N/A				
lace: Warszawa ate of issue: 2023-04-0	1/6					

Signature:\_

Andrzej Cegliński

Quality Assurance and Rfegulatory Compliance Manager

TEMPLATE NO	EA7-SOP QARC-020 TEM-04	REVISION LEVEL	1
-------------	-------------------------	----------------	---